

JUL 14 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130133</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Garrett R. Willis</u>	ID# <u>5560</u> Date <u>5-26-11</u>
A Agency <u>AST</u> Phone # <u>783-0972</u>	
Instrument Location <u>Girdwood</u>	
B Alco S/N <u>633421</u> Target Value <u>0.086</u> High Pressure _____	
Alco Test Values <u>0.082</u> <u>0.082</u> 1st Alco 2nd Alco	
Signature <u>Garrett R. Willis</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 2010. NOB

Nikki Roth (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



JUL 14 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130133

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

MAY 26, 2011

OPERATOR'S NAME:

WILLIS/GARRETT/RO

OPERATOR'S NUMBER: 5560

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: 6040

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .006

ALCO S/N: 633421

--- BREATH ANALYSIS ---

.006 ADJUSTED FOR 30.12 in
ALCO TARGET .006 12:46
BLANK TEST .000 12:47
INTERNAL STANDARD VERIFIED 12:47
ALCO TV 30.14 in .002 12:48
BLANK TEST .000 12:48
SUBJECT SAMPLE .000 12:49
BLANK TEST .000 12:50
ALCO TV 30.14 in .002 12:50
BLANK TEST .000 12:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

MAY 26, 2011

TIME 13:50

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 4c
BAROMETER: 30.19 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

HIJKL MNOPQRSTUVWXYZ [^] _ ` abcdefghijklmno
pqrstuvwxyz{|}~